

UN 38.3 Test Summary Report

[a] Name of cell, battery or product manufacturer, as applicable	[b] Cell, battery, or product manufacturer's contact information to include address, phone number, email address and website for more information;		
Item Number : LTC-3PN Item Name : Cell Item Description : 3.6V / 0.35Ah Same Type Part - LTC-3PN - LTC-3PN-S2 - LTC-3PN-S4 - LTC-3PN-S10 - LTC-3PN-SMS2	Xenoenergy Co., Ltd. 70-7, Muha-ro, Hwaseong-city, Kyounggi-do Republic of Korea, 18729 Tel : +82 31 355 3511 E-mail : global@xenoenergy.com Website : www.xenoenergy.com		
[c] Name of the test laboratory to include address, phone number, email address	[d] A unique test report [e]Date of test report;		
and website for more information;	identification number;		,
	Xeno Q101203	December. 03, 20	010
Xenoenergy Co., Ltd.			
70-7, Muha-ro, Hwaseong-city, Kyounggi-do Republic of Korea, 18729	[g] List of tests conducted and results (i.e., pass/fail);		
Tel : +82 31 355 3511 E-mail : R&D@xenoenergy.com Website :www.xenoenergy.com	☑ Test T.1 : Altitude simulation		T1 Test Pass
	✓ Test T.2 : Thermal test		T2 Test Pass
[f] Description of cell or battery to include at a minimum: Lithium ion or lithium metal cell or battery;Mass; Watt-hour rating,	✓ Test T.3 : Vibration		T3 Test Pass
or lithium content; Physical description of the cell/battery; and Model numbers.	✓ Test T.4 : Shock		T4 Test Pass
Cell used in implantable medical devices,	✓ Test T.5 : External short circuit		T5 Test Pass
such as : Pacemaker(IPG),Neuro-Stimulators or Implantable Defibrillators(ICD)	✓ Test T.6 : Impact/Crush(cell only test)		T6 Test Pass
	✓ Test T.7 : Overcharge(N.A for Li-metal only)		Not Applicable
(i) Lithium metal (ii) Mass : max. 5g (iii) Lithium content : max 0.2g	✓ Test T.8 : Forced discharge(cell only test) T8		T8 Test Pass
(iv) ELC or W/h rating : 3.6V / 0.35Ah (v) Model number: LTC-3PN	Testing additional comments :		
	No additional information		
[h] Reference to assembled battery testing requirements, if applicable (i.e. 38.3.3 (f) and 38.3.3 (g));	[i] Reference to the revised edition of the Manual of Tests and Criteria used and to amendments thereto, if any;		
Not Applicable.	UN Manual of Tests and Criteria, Part III, sub-section 38.3. sixth revised edition		
 Signature with name and title of signatory as an indication of the validity of information provided. 	Date this doucment was generated.		
Assistant Managing Director Director of R&D Center JR Kim	Oct. 21, 2019		